

**E-20 Surgical Consumables (Technical Queries)**

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
1	1	Absorbable unidirectional knotless tissue control device, symmetric anchoring pattern with 5 anchors/cm , triclosan coated polydioxanone with fixation tab , size 1-0, 36mm 1/2 circle TAPER POINT needle, 45 cm	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend as Absorbable unidirectional knotless tissue control device,symmetric anchoring pattern with 5 anchors/cm,polydioxanone with fixation tab,size 1-0, 36 mm, 1/2 circle taper point needle, 45 cm	No changes
			M/s Earthnine Naturals LLP	We request removal of Triclosan in Barb Knotless Sutures / Monofilament Sutures and PDS Sutures (Sr. No. 1, 2, 8, 9). Additionally, kindly verify and align the PDS Antibacterial Suture serial number. This amendment will ensure fair competition and allow participation from multiple bidders.	No changes
			M/s Meril Endo Surgery Pvt Ltd.	Remove triclosan Coated	No changes
			M/s Healthium Medtech Limited	Please ammend Absorbable Surgical Needled-Suture (Synthetic) Unidirectional Knotless Tissue Closure/ Control Device Monofilament Polydioxanone Size: 1-0, 1/2 Circle Taper point, 36-37mm, 45 cm	No changes
			M/s TI Medical Private Limited	Please ammend 37MM needle, Loop at the end, No Triclosan coating & Symmetric Anchoring Pattern.	No changes

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
2	2	Absorbable unidirectional knotless tissue control device, symmetric anchoring pattern with 5 anchors/cm, triclosan coated polydioxanone with fixation tab, size 2-0, 36mm 1/2 circle TAPER POINT needle, 45 cm	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend as Absorbable unidirectional knotless tissue control device,symmetric anchoring patternwith 5 anchors/cm, polydioxanone with fixation tab,size 2-0, 36 mm, 1/2 circle taperpoint needle, 45 cm	No changes
			M/s Earthnine Naturals LLP	We request removal of Triclosan in Barb Knotless Sutures / Monofilament Sutures and PDS Sutures (Sr. No. 1, 2, 8, 9). Additionally, kindly verify and align the PDS Antibacterial Suture serial number. This amendment will ensure fair competition and allow participation from multiple bidders.	No changes
			M/s Meril Endo Surgery Pvt Ltd.	Remove triclosanCoated	No changes
			M/s Healthium Medtech Limited	Please ammend Absorbable Surgical needled Suture (Synthetic) Knotless Tissue- Closure/ Control Device Monofilament Polydioxanone - Barbed Suture with a Triangular Stopper Size: 2-0, 1/2 Circle Taper Point, 45 cm, 36-37 mm Unidirectional	No changes
			M/s TI Medical Private Limited	Please ammend 37MM needle, Loop at the end, No Triclosan coating & Symmetric Anchoring Pattern	No changes
3	3	Two-point absorbable Fixation device for Lap hernia repair. Trigger squeeze mechanical device with straight cannula, pre-loaded with 25 absorbable straps which are made of a blend of polydioxanone dyed with D&C Violet No. 2 and L(-)-lactide/Glycolide copolymer. Inserted length of strap should be 6.7 mm	M/s Meril Endo Surgery Pvt Ltd.	Remove two point strap	No changes
			M/s TI Medical Private Limited	Please ammend Helical shape tackers 7mm height, without 2 point fixation	No changes

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
4	4	Tissue separating multilayer mesh for intra-abdominal open and laparoscopic ventral hernia repair. Soft polypropylene mesh clubbed with polydioxanone layer and has an oxidized regenerated cellulose on visceral side.(Size-15cmx15cm)	M/s Meril Endo Surgery Pvt Ltd.	Kindly consider our polypropylene +polylactide-caprolactonetissue separating dual layer mesh	No changes
5	5	Polypropylene Mesh Size-7.5cmx15cm with 1 Poliglecaprone25 and 2 Polypropylene suture 2-0,	M/s Advanced Medtech Solutions Pvt Ltd	Please clarify the exact requirement	Kit comes with suture as per specification
			M/s Earthnine Naturals LLP	the item pertains to a Hernia Kit. We request you to kindly amend the same accordingly	Kit comes with suture as per specification
6	8	Triclosan Antibacterial coated Monofilament Poliglecaprone 25 suture, length 70 cm ,size 3-0 with 3/8 Circle reverse cutting , 26mm needle	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend Monofilament Poliglecaprone 25 suture, length 70 cm, size 3-0 with 3/8 circle reverse cutting,26 mm needle	No changes
			M/s Earthnine Naturals LLP	We request removal of Triclosan in Barb Knotless Sutures / Monofilament Sutures and PDS Sutures (Sr. No. 1, 2, 8, 9). Additionally, kindly verify and align the PDS Antibacterial Suture serial number. This amendment will ensure fair competition and allow participation from multiple bidders.	No changes
			M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification	No changes
			M/s TI Medical Private Limited	Triclosan coating to be removed	No changes

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
6	9	Triclosan Antibacterial coated Monofilament Poliglecaprone 25 suture, length 70 cm ,size 4-0 with 3/8 Circle reverse cutting , 16mm needle	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend Monofilament Poliglecaprone 25 suture, length 70 cm, size 4-0 with 3/8 circle reverse cutting, 16 mm needle	No changes
			M/s Earthnine Naturals LLP	We request removal of Triclosan in Barb Knotless Sutures / Monofilament Sutures and PDS Sutures (Sr. No. 1, 2, 8, 9). Additionally, kindly verify and align the PDS Antibacterial Suture serial number. This amendment will ensure fair competition and allow participation from multiple bidders.	No changes
			M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification	No changes
			M/s TI Medical Private Limited	Triclosan coating to be removed	No changes
7	10	20 Large titanium clips for ligating Large Size vessels with Shaft length of 33.7 cm with Ratchet closure, Proprietary lateral and transverse grooves on inner clip surface enhances on-vessel clip security. Anti-clip drop, Complete clip closure designed to provide proximal to distal closure with minimal gap.	M/s Advanced Medical Solutions	Please ammend as 20 Large Titanium Clips For Ligating Large Size Vessels With Shaft Length 33 cm to 35 cm With endo applicator , Lateral And Transverse Grooves On Inner Clip Surface Enhances On-Vessel Clip Security. Anti-Clip Drop, Complete Clip Closure Designed To Provide Proximal To Distal Closure With Minimal Gap.	No changes
8	11	Non absorable polybutylate coated braided polyester 1/2 circle taper cut V-55 needle 55mm green braided 90cm size 5, 4 SUTURE PER FOIL	M/s TI Medical Private Limited	Suture length to be ammended as 75cm	No changes
9	12	Triclosan Antibacterial coated Violet Monofilament Polydioxanone suture with 70 cm size 2-0 with 1/2 Circle Taper Point SH, 26mm needle USFDA approved.	M/s Meril Endo Surgery Pvt Ltd.	Provide Needle 26mm-31mm	Deviation upto 2mm will be acceptable

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
10	13	Triclosan Antibacterial coated Violet Monofilament Polydioxanone suture with 70 cm size 3-0 with 1/2 Circle Taper Point RB-1, 17mm needle USFDA approved.	M/s Meril Endo Surgery Pvt Ltd.	Provide Needle 17mm-20mm	Deviation upto 2mm will be acceptable
			M/s TI Medical Private Limited	Triclosan coating & USFDA Approved to be removed & criteria be revised to: "USFDA / CE / CDSCO / BIS (as applicable)	No change
11	14	Triclosan Antibacterial coated Violet Monofilament Polydioxanone suture with 70 cm size 1 with 1/2 Circle Taper Point SH, 48mm needle USFDA approved.	M/s TI Medical Private Limited	Triclosan coating & USFDA Approved to be removed & criteria be revised to: "USFDA / CE / CDSCO / BIS (as applicable)	No change
12	15	Non absorbable polybutylate coated braided polyester 1/2 circle taper cut 26mm needle green braided 90cm size 2-0	M/s Meril Endo Surgery Pvt Ltd.	Add Double Armed Needle	No change
13	16	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle round body 40mm heavy Needle violet brided coated - 90cm, size-1. with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 1/2 Circle Round Body 40mm Heavy Needle Violet Braided Coated-90cm, Size-1	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change
14	17	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 40mm OS Needle violet brided coated - 90cm, size-1. with irgacare MP impregnated.	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 1/2 Circle Reverse Cutting 40mm OS/ OR Needle Violet Braided Coated - 90cm, Size-1	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
15	18	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle round body 40mm heavy Needle violet bridged coated - 90cm, size-1-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 1/2 Circle Round Body 40mm Needle Violet Braided Coated - 90cm, Size-1-0.	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change
16	19	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle reverse cutting 40mm Needle violet bridged coated - 90cm, size-2-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 1/2 Circle Reverse Cutting 40mm Needle Violet Braided Coated- 90cm, Size-2-0	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
17	20	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 circle ROUND BODY 30mm Needle violet brided coated - 90cm, size-2-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 1/2 Circle Round Body 30mm Needle Violet Braided Coated - 90cm, Size-2-0	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change
18	21	Antibacterial Absorbable Surgical suture Polyglactin 910 3/8 circle REVERSE cutting 24mm Needle violet brided coated 70cm, size-3-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 3/8 Circle Reverse Cutting 24mm Needle Violet Brided Coated - 70cm, Size-3-0	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change
19	22	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 CIRCLE TAPER POINT MULTIPASS 20MM Needle violet brided coated - 70cm, size-3-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Meril Endo Surgery Pvt Ltd.	suture length 70-90cm	No changes
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated & Please ammend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 1/2 Circle Taper Point/ Round Body 20mm Needle Violet Braided Coated - 70cm, Size-3-0	No change
			M/s TI Medical Private Limited	Request Irgacare MP Impregnated to be removed	No change

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
20	23	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 CIRCLE ROUND BODY 20MM Needle violet bridged coated - 70cm, size-4-0.with Triclosan coating	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
21	24	Absorbable Surgical suture Polyglactin 910 3/8 CIRCLR CUTTING PC3 16MM Needle violet bridged coated - 45cm, size-4-0.irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s Healthium Medtech Limited	Request you to delete Irgacare Mp Impregnated Pc3 and please amend Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910 3/8 Circle Cutting 16mm Needle Violet Braided Coated - 45cm, Size-4-0	No change
22	25	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 CIRCLE ROUND BODY 17MM Needle violet bridged coated - 45cm, size-5-0.with Triclosan coating	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
			M/s TI Medical Private Limited	Standard Needle size is 16mm	No changes
23	26	Absorbable Surgical suture Polyglactin 910 1/4 CIRCLE SPATULATED MICRO POINT DOUBLE 8MM Needle violet bridged coated - 45cm, size-6-0.	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change
24	27	Antibacterial Absorbable Surgical suture Polyglactin 910 3/8 CIRCLE REVERSE CUTTING 22MM Needle violet bridged coated - 45cm, size-3-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No change

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
25	29	Size 12mm X 100MM, lowest peak insertion force with 12mm diameter and 100 mm length with low insertion force , Abdominal wall retention & Integrated thread design with reverse Christmas tree grooves on bladeless trocar cannulas. It should have open upper housing for specimen removal with taking out trocar sleeve from abdominal space Duck bill seal which maintains pneumoperitoneum and Durable universal seal which accommodates a wide range of instrument sizes (4.7mm to 12.9mm)	M/s Meril Endo Surgery Pvt Ltd.	Provide Trocar Length 95mm-100	No change
26	35	Non Absorbable, Monofilament Polyamide BLACK Sterile Surgical Suture length 38cm ,size-4-0, 3/8 Circle, Reverse cutting, Needle, 11mm	M/s Meril Endo Surgery Pvt Ltd.	Provide Needle10mm-11mm	Acceptable
			M/s TI Medical Private Limited	Standard Needle size is 10mm	Acceptable

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
27	36	Non Absorbable, blue Monofilament, Polypropylene Sterile Surgical Suture length , 60cm,size 7-0, 3/8 Circle, BV175-6 TUNGSTEN RHENIUM ALLOY, Double Needle, 8mm	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend Non-absorbable blue monofilament polypropylene sterile surgical suture, length 60 cm, size 7-0, 3/8 circle, BV175-6,double needle, 8 mm	No change
			M/s Earthnine Naturals LLP	Tungsten Rhenium Alloy in Monofilament Sutures restricts participation to a single manufacturer. We request amendment of this specification	No changes
			M/s Meril Endo Surgery Pvt Ltd.	Remove Tungsten Rhenium Alloy	No change
			M/s Healthium Medtech Limited	please delete rhenium tungsten alloy, Bv175-6 and Please amend the Item Name: Non Absorbable, Blue Monofilament, Polypropylene Sterile Surgical Suture Length, 60cm,Size 7-0, 3/8 Circle, Taper Point, Double Needle, 8mm	No change
28	37	Non Absorbable, blue Monofilament, Polypropylene Sterile Surgical Suture length , 60cm,size 8-0, 3/8 Circle, BV175-6 TUNGSTEN RHENIUM ALLOY, Double Needle, 8mm	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend Non-absorbable bluemonofilament polypropylene sterile surgical suture, length 60 cm, size 8-0, 3/8 circle, BV175-6,double needle, 8 mm	No change
			M/s Earthnine Naturals LLP	Tungsten Rhenium Alloy in Monofilament Sutures restricts participation to a single manufacturer. We request amendment of this specification	No change
			M/s Meril Endo Surgery Pvt Ltd.	Remove Tungsten Rhenium Alloy	No change
			M/s Healthium Medtech Limited	please delete rhenium tungsten alloy, Bv175-6 Please amend the Item Name: Non Absorbable, Blue Monofilament, Polypropylene Sterile Surgical Suture Length, 60cm, Size 8-0, 3/8 Circle Taper Point Double Needle, 8mm	No change

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
29	38	<p>Size 5 mm X 100MM, lowest peak insertion force with 5mm diameter and 100 mm length with low insertion force , Abdominal wall retention &amp; Integrated thread design with reverse Christmas tree grooves on bladeless trocar cannulas. It should have open upper housing for specimen removal with taking out trocar sleeve from abdominal space Duck bill seal which maintain pneumoperitoneum and Durable universal seal which accommodates a wide range of instrument sizes.</p>	M/s Meril Endo Surgery Pvt Ltd.	Provide Trocar Length 95mm-100	No change

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
30	39	Non Absorbable, blue Monofilament, Polypropylene Sterile Surgical Suture length , 60cm,size 7-0, 3/8 Circle, BV175-8 ETHALLOY (220 Microns), Double Needle, 9.3mm	M/s Meril Endo Surgery Pvt Ltd.	Provide suture length 60cm-75cm	Deviation upto 5cm is acceptable
			M/s Healthium Medtech Limited	please delete Ethalloy (220 Microns), BV175-8. Please amend the Item Name: Non Absorbable, Blue monofilament , ploypropylene sterile surgical suture length, 60cm, size, 7-0, 3/8 circle, round body CC, Double needle, 9.3 mm	No changes
			M/s TI Medical Private Limited	Standard Needle size is 9mm	No changes
31	40	ABSORABLE HEMOSTATIC MATRIX ABSORB COMPLETELY WITHIN 4 TO 6 WEEKS AVAILABLE BOTH POWDER AND SPONG FORM INDICATED TO USE WITH THROMBIN OR STERILE SALINE	M/s Baxter India Pvt Ltd.	Absorbable Hemostatic Matrix inPre-filled syringe of Gelatin alongwith Thrombin & Sodium Chloride Solution for reconstitution withsize 5ml to 10ml & malleable applicator tip is included.	Acceptable
			M/s Aegis LifeSciences	Kindly clarify, whether the requirement is for matrix form or powder form,	Matrix form with thrombin required
32	42	Absorbable unidirectional knotless tissue control device, spiral anchoring pattern with anchors at 15 degrees and 20 anchors/inch with an adjustable fixation loop at one end of the suture, triclosan coated polydioxanone, tensile strength 2-0, 36mm 1/2 circle reverse cutting needle, 15 cm	M/s Meril Endo Surgery Pvt Ltd.	Remove triclosan Coated	No changes
33	43	Antibacterial Absorbable Surgical suture Polyglactin 910 1/2 CIRCLE REVERSE CUTTING 22MM Needle violet bridged coated - 35cm, size-1.with Triclosan coating	M/s Healthium Medtech Limited	Please ammend AntibacterialAbsorbable Surgical SuturePolyglactin 910 1/2 Circle Reverse Cutting 22-23mm Needle Violet Braided Coated - 35cm, Size-1. With Triclosan Coating	Deviation of 2mm in neddle is acceptable

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
34	44	Non Absorbable, blue Monofilament, Polypropylene Sterile Surgical Suture length 70cm ,size-3-0, 1/2 Circle, TAPER POINT MULTIPASS, Needle, 26mm	M/s Healthium Medtech Limited	Plse delete word " Multipass" and ammend Non Absorbable, Blue Monofilament, Polypropylene Sterile Surgical Suture Length 70cm ,Size-3-0, 1/2 Circle, Taper Point, Needle, 25-26mm	No changes
			M/s TI Medical Private Limited	Standard Needle size is 25mm	No changes
35	45	Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA size 8"x4" inch	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend US FDA Including ISO certification aligns with QMS standards	Acceptable
			M/s Earthnine Naturals LLP	we request inclusion of ISO / USFDA certification. ISO certification aligned with QMS standards also supports the "Make in India" initiative	Acceptable
			M/s Meril Endo Surgery Pvt Ltd.	Instead of Non-woven provide standard ORC	No changes
			M/s Aegis LifeSciences	If the requirement is specifically for the Snow (non-woven) variant, it is typically available in sizes such as 1"x2", 2"x4", and 4"x4" We kindly request clarification on the required variant and size combination	Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA size 4"x4" inch
M/s Advanced Medtech Solutions Pvt Ltd	Please ammend US FDA Including ISO certification aligns with QMS standards				
36	46	Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA size 2"x3" inch	M/s Earthnine Naturals LLP	we request inclusion of ISO / USFDA certification. ISO certification aligned with QMS standards also supports the "Make in India" initiative	Oxidized Regenerated Cellulose based Topical Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA size 4"x4" inch
			M/s Meril Endo Surgery Pvt Ltd.	Instead of Non-woven provide standard ORC	
			M/s Aegis LifeSciences	We request clarification on the required specification	
			M/s Meril Endo Surgery Pvt Ltd.	Conside single suture per foil	Oxidized Regenerated Cellulose based Topical Absorbable Hemostat

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
37	48	polyester 1/2 circle taper cut v-59 needle 44.8mm green braided 75cm size 2, 4 SUTURE PER FOIL	M/s Healthium Medtech Limited	Please ammend Non Absorbable Polybutylate / Silicone Coated Braided Polyester 1/2 Circle Taper end Cutting Heavy Needle 44.8- 45 mm Green Braided 75cm Size 2, 4 Suture Per Foil	Absorbable Hemostat, Structured Non-Woven material, with Bactericidal property. Ease-of-use in both open and minimally invasive procedures. Approved by US FDA size 2"x4" inch
38	49	Absorbable unidirectional knotless tissue control device, symmetric anchoring pattern with 5 anchors/cm , triclosan coated polydioxanone with fixation tab , size 1-0, 36mm 1/2 circle TAPER POINT needle, 45 cm	M/s Meril Endo Surgery Pvt Ltd.	Remove triclosan Coated	
			M/s Healthium Medtech Limited	Please ammend Absorbable Surgical Needled-Suture (Synthetic) Unidirectional Knotless Tissue Closure/ Control Device Monofilament Polydioxanone Size 1-0, 36-37mm 1/2 Circle Taper Point Needle, 45 Cm	No changes
			M/s TI Medical Private Limited	Please ammend 37MM needle, Loop at the end, No Triclosan coating & Symmetric Anchoring Pattern	No changes
39	54	Advanced bipolar tissue sealer with 5mm shaft diameter, 24 mm Jaw length, 21.8 mm cut length and 13.4 mm Jaw aperture designed for use in laparoscopic surgical procedures with curved and tapered tip and uses an advanced algorithm for intelligent and efficient energy delivery with shaft length 25 cm. Device to have intuitive design with separate Seal and Cut button and 360 degree continuous shaft rotation.	M/s Johnson and Johnson Private Limited (Ethicon)	Please ammend Advanced Bipolar Tissue Sealer With 5mm Shaft Diameter, 24 Mm Jaw Length, 21.8 Mm Cut Length And 13.4 Mm Jaw Aperture Designed For Use In Laparoscopic Surgical Procedures With Curved And Tapered Tip And Uses An Advanced Algorithm For Intelligent And Efficient Energy Delivery With Shaft Length 25 and 37 Cm. Device To Have Intuitive Design With Separate Seal And Cut Button And 360 Degree Continuous Shaft Rotation.	Acceptable
40	55	Braided Coated Polyester Suture 5-0, 1/2 circle spatulated micro point Double Needle 8mm Length 45cm	M/s Meril Endo Surgery Pvt Ltd.	Instead of 1/2circle needle consider 1/4circle needle	No changes
41	56	Polypropylene Mesh Size- 6cmx11cm with 1 Poliglecaprone25 and 2 Polypropylene suture 2-0,	M/s Advanced Medtech Solutions Pvt Ltd	Please clarify the exact requirement	Hernia kit with that mesh size

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
42	57	Absorbable surgical Polydioxanone sterilized surgical suture 3/8 Circle, TAPER POINT, 5-0, 17mm, monofilament, 70cm.	M/s Healthium Medtech Limited	Please ammend Absorbable Surgical Polydioxanone Sterilized Surgical Suture 3/8 Circle, Taper Point/ Round Body, 5-0, 13-17mm, Monofilament, 70cm	Deviation of 2mm in needle is acceptable
43	69	Size 7mm, Laparoscopic ultrasonic ergonomic curved coagulating shears 36 cm length and 5 mm shaft diameter with 360 degree rotation with adaptive tissue technology and capable of sealing 7 mm vessels only with pure ultrasonic energy. It must be compatible with the Ultrasonic generator GEN11 and handpiece	M/s Meril Endo Surgery Pvt Ltd.	Change 7mm vessel sealing to 5mm	No changes
44	71	Polyamide Suture 4-0, 3/8 circle cutting 19mm Length 70cm	M/s Healthium Medtech Limited	Please ammend Polyamide Suture 4-0, 3/8 Circle Cutting 16-19mm Length 70cm	Acceptable
45	73	Polyamide Suture 5-0, 3/8 circle cutting slim blad 16mm Length 70cm	M/s Healthium Medtech Limited	Please ammend Polyamide Suture 5-0, 3/8 Circle Cutting Slim Blade 15-16mm Length 70cm	Acceptable
46	85	Polypropylene Suture 5-0, 3/8 circle reverse cutting 13mm Length 70cm	M/s Healthium Medtech Limited	Please ammend Polypropylene Suture 5-0, 3/8 Circle Reverse Cutting 12-13mm Length 70cm	Acceptable
47	86	Polypropylene Suture 6-0, 3/8 circle round body DN 9.3mm Length 60cm	M/s Healthium Medtech Limited	Please ammend the name as Polypropylene Suture 6-0, 3/8 Circle Round Body/ Taper point DA 9-9.3mm Length 60cm	Acceptable
48	88	Polypropylene Suture 6-0, 3/8 circle cutting 16mm Length70cm	M/s Meril Endo Surgery Pvt Ltd.	Provide Needle13mm-16mm	Acceptable
			M/s TI Medical Private Limited	Standard Needle size is 13mm	Acceptable
49	98	Blue Monofilament Polypropylene, size 2-0 ,1/2 Circle Taper Cut Double Armed, 26 mm Needle, 90 cm length	M/s HLL Lifecare Ltd	Request you to put needle size between 25-30	Acceptable
			M/s TI Medical Private Limited	Standard Needle size is 25mm	Acceptable
50	99	Blue Monofilament Polypropylene, Size 4-0, 1/2 Circle Taper Point Double Armed, 26 mm Needle, 90 cm Length	M/s HLL Lifecare Ltd	Request you to put needle size between 25-30	Acceptable

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
51	100	Blue Monofilament Polypropylene, Size 4-0, 1/2 Circle Taper Point Double Armed, 16 mm Needle, 90 cm Length	M/s HLL Lifecare Ltd	Request you to put needle size between 15-20	Acceptable
52	108	Electrode Blue 60 cm Pacing Wire, Pediatric, 18 mm 1/2 Circle Taper Cut & 63 mm Straight Cutting Edge, Breakway Needle	M/s Meril Endo Surgery Pvt Ltd.	Kindly consider our Electrode orange 60 Cm Pacing Wire, Pediatric, 17 Mm1/2 Circle Taper Cut & 60 Mm Straight Cutting Edge, Breakway Needle	Acceptable
53	109	Electrode Blue 60 cm Pacing Wire, Adult, 26 mm 1/2 Circle Taper Cut & 63 mm <b>Straight Cutting Edge</b> , Breakway Needle	M/s Meril Endo Surgery Pvt Ltd.	Kindly consider our Electrode orange 80 Cm Pacing Wire, Adult, 26 Mm1/2 Circle Taper Cut & 60 Mm Straight Cutting Edge, Breakway Needle	Acceptable
54	117	Monofilament Stainless Steel 4 x 45 cm, Size 5, 316 L 1/2 Circle Blunt Point Round Bodied, 45mm Needle, 45cm Length	M/s Healthium Medtech Limited	Please ammend the name as Monofilament Stainless Steel 4 X 45 Cm, Size 5,316 L 1/2 Circle Blunt Point Round Bodied, 44-45mm Needle, 45cm Length	Acceptable
			M/s TI Medical Private Limited	Standard Steel suture length is 45cms	Acceptable
55	118	Monofilament Stainless Steel 4 x 45 cm, Size 5, 316 L 1/2 Circle Blunt Point Round Bodied, 45mm Needle, 48cm Length	M/s Healthium Medtech Limited	Please ammend the name as Monofilament Stainless Steel 4 X 45 Cm ,Size 5,316 L 1/2 Circle Blunt Point Round Bodied, 44-45mm Needle, 48cm Length	Acceptable
			M/s TI Medical Private Limited	Standard Steel suture length is 45cms	Acceptable
56	119	Monofilament Stainless Steel 4 x 45 cm, Size 6, 316 L 1/2 Circle Blunt Point Round Bodied, 45mm Needle, 48cm Length	M/s Healthium Medtech Limited	Please ammend the name as Monofilament Stainless Steel 4 X 45 Cm, Size 6,316 L 1/2 Circle Blunt Point Round Bodied, 44-45mm Needle, 48cm Length	Acceptable
57	121	Non Absorbable, blue Monofilament, Polypropylene Sterile Surgical Suture, length - 60cm, size 8-0, 3/8 Circle, BV175-6 TUNGSTEN RHENIUM ALLOY, Double Needle, 8mm	M/s Advanced Medtech Solutions Pvt Ltd	Please ammend Non-absorbable blue monofilament polypropylene sterile surgical suture, length 60 cm, size 8-0, 3/8 circle, BV175-6, double needle, 8 mm	No changes
			M/s Earthnine Naturals LLP	Tungsten Rhenium Alloy in Monofilament Sutures restricts participation to a single manufacturer. We request amendment of this specification	No changes
			M/s Meril Endo Surgery Pvt Ltd.	Remove Tungsten Rhenium Alloy	No changes

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
58	122	Non Absorbable, blue Monofilament, Polypropylene Sterile Surgical Suture, length - 60cm,size 7-0, 3/8 Circle, BV175-8 ETHALLOY (220 Microns), Double Needle, 9.3mm	M/s Meril Endo Surgery Pvt Ltd.	Kindly consider Non Absorbable, Blue Monofilament, Polypropylene Sterile Surgical Suture, Length -60cm,Size 7-0, 3/8Circle round body,175-7, Double Needle, 8mm	No changes
59	123	Antibacterial Absorbable Surgical suture Polyglactin 910, 1/2 circle ROUND BODY 30mm Needle violet brided coated 90cm, size-2-0.with irgacare MP impregnated	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	No changes
			M/s Healthium Medtech Limited	please delete Irgacare Mp Impregnated and please ammend name as Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910, 1/2 Circle Round Body 30mm Needle Violet Braided Coated - 90cm, Size-2-0.	No changes
			M/s TI Medical Private Limited	Irgacare MP Impregnated to be removed	No changes
60	124	Antibacterial Absorbable Surgical suture Polyglactin 910, 1/2 circle round body 40mm heavy Needle violet brided coated - 90cm, size-1-0.with irgacare MP impregnated	M/s Healthium Medtech Limited	please delete Irgacare Mp Impregnated and please ammend name as Antibacterial with Triclosan Absorbable Surgical Suture Polyglactin 910, 1/2 Circle Round Body 40mm Heavy Needle Violet Braided Coated - 90cm, Size-1-0.	No changes
			M/s TI Medical Private Limited	Irgacare MP Impregnated to be removed	No changes
61	125	Absorbable Hemostat (ORC) - 1x2	M/s Aegis LifeSciences	Kindly confirm the required variant, as these sizes are available in both Fibrillar and Snow (non-woven) forms	Orc 2 x 3 with clinical efficacy supported by scientific data rct
62	126	Absorbable Hemostat (ORC) - 2x4	M/s Aegis LifeSciences	The specified size is 2x4 inches; however, the variant has not been mentioned. We offer this size in two variants, namely Fibril and Snow (Non-woven). We request you to kindly clarify the required variant	Orc 4 x 8 with clinical efficacy supported by scientific data rct
63	127	Chromic Catgut Size 3-0, 20mm, 1/2 Circle, Round Body	M/s HLL Lifecare Ltd	please consider Or additionally put the same products without Antibacterial specification.	Acceptable

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
64	129	SYNTHETIC ABSOABLE COATED ANTIBACTERIAL SUTURES (POLYGLACTIN 910 WITH TRICLOSAN) 3/8 Circle Cutting Ethiprime 16mm VICRYL PLUS Violet Braided Antibacterial 45 cm SIZE 5-0	M/s TI Medical Private Limited	Brand name & Ethiprime needle to be removed.	SYNTHETIC ABSOABLE COATED ANTIBACTERIAL SUTURES (POLYGLACTIN 910 WITH TRICLOSAN) 3/8 Circle Cutting prime 16mm Violet Braided Antibacterial 45 cm SIZE 5-0
65	132	SYNTHETIC ABSOABLE (POLYGLACTIN 910) COATED, 1/2 Circle Reverse Cutting (Heavy) Portt 22 mm, Violet Braided 45 cm SIZE 0	M/s Meril Endo Surgery Pvt Ltd.	Provide Needle Dimension 22mm-13mm	Deviation of 2mm in neddle is acceptable
			M/s Healthium Medtech Limited	please ammend name as Synthetic Absoable (Polyglactin 910) Coated,1/2 Circle Reverse Cutting 22-23 Mm, Violet Braided 45 Cm Size 0	Deviation of 2mm in neddle is acceptable
			M/s TI Medical Private Limited	Standard Needle size is 23mm	Deviation of 2mm in neddle is acceptable
66	134	SYNTHETIC ABSOABLE MONOCRYL (POLIGLECAPRONE 25) 3/8 Circle Reverse Cutting ETHIPRIME 16 mm, Undyed Monofilament 70 cm SIZE 4-0	M/s Healthium Medtech Limited	Please delete the word "MONOCRYL, Ethiprime" and Please ammend Item Name:Synthetic Absorbable (Poliglecaprone 25) 3/8Circle Reverse Cutting 16 Mm, Undyed Monofilament 70 Cm Size 4-0	SYNTHETIC ABSOABLE (POLIGLECAPRONE 25) 3/8 Circle Reverse Cutting PRIME 16 mm, Undyed Monofilament 70 cm SIZE 4-0

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
67	i	2.24 Similarly for products other than medicines i.e. (For surgical and consumables) WHO GMP/CE /ISO/ BIS whichever applicable may be attached along with authorization letter .NABL accreditation wherever required, USFDA/PDWA(Japan) wherever required.	M/s Earthnine Naturals LLP	ISO certification has been accepted as sufficient quality compliance, and USFDA certification has not been mandatory	No changes
			M/s Healthium Medtech Limited	Please consider delivery period as minimum 60-75 days instead of 45 days.	
			M/s TI Medical Private Limited	We request that certification criteria be revised to: "USFDA / CE / CDSCO / BIS (as applicable)	
			M/s Aegis LifeSciences	we humbly request your good office to consider acceptance of products compliant with Indian regulatory standards, including CDSCO approval, ISO certification, and CE (MDR), in lieu of USFDA certification	
			M/s Johnson & Johnson Private Limited	THE SAID CLAUSE IS NOT MANDATORY . Clarification to be issued in this regard.	
			M/s Mediearth lifecare Pvt Ltd	we request clarification that this clause should not be treated as mandatory beyond CDSCO/SLA licensing requirements	

Sr. No.	Tender Sr No	Name of the Item	Name of the Bidder	Query / Clarification	Expert Remark
68	ii	Regarding Sample Submission	M/s Healthium Medtech Limited	Please confirm sample qty and due date for submission of sample	Three Pieces of each items Should be submitted at DMER Office before date 15/06/2026 5.00pm
			M/s Aegis LifeSciences	Please confirm the number of samples required for submission, and the timeline for submission of samples.	
70	iii	2.18 Valid WHO & DCGI Certificate of manufacturer, NABL Test report is compulsory	M/s Johnson & Johnson Private Limited	DCGI Certificate of manufacturer and QMS Certificate of the Manufacturer to be accepted.	Acceptable
			M/s Advanced Medical Solutions	We request deletion of the term "WHO" from this clause	
			M/s Mediearth lifecare Pvt Ltd	As the tendered products fall under the Medical Devices Rules, 2017, WHO GMP certification is not applicable. We request deletion of the term "WHO" from this clause.	
71	iv	2.20 Valid WHO & DCGI Certificate with product list / COPP (product-wise). Validity must be mentioned on the certificate; otherwise, bidder COPP will not be considered.	M/s Johnson & Johnson Private Limited	As the said tendered products are comes under Medical Devices Rules 2017, WHO GMP certificate is not applicable, therefore DCGI Certificate of manufacturer and QMS Certificate of the Manufacturer to be accepted	Medical Device Licese is applicable WHO GMP Certificte is deleted However QMS Certificate is Preferable
			M/s Advanced Medical Solutions	We request deletion of the term "WHO" from this clause and Acceptance of DCGI Certificate along with QMS Certificate of the manufacturer.	
			M/s Mediearth lifecare Pvt Ltd	We request Deletion of the term "WHO", and Acceptance of DCGI Certificate along with QMS Certificate of the manufacturer.	
72	v	Suture length and needle sizes	M/s TI Medical Private Limited	Suture Length and needle variation acceptance range to be mentioned i.e plus minus ( $\pm$ ) 5mm in needle and 5cm in suture length.	No changes



Director


Medical Education & Research  
Mumbai 400001

**E-20 Surgical Consumables (Admin Queries)**

Sr. No.	Tender Clause and No	Name of the Bidder	Query / Clarification	Department Remark and Signature
1	2.6 The bidder should have a minimum annual turnover of Rs.13 crore per year shall be as indicated in past Three financial years i.e. 2022-23, 2023-24 & 2024-25 As per Annexure-3 certified by the Chartered Accountant. Joint venture between Manufacturer and exclusive distributor can be Considered	M/s Earthnine Naturals LLP	We request you to kindly consider an average annual turnover of ₹13 Crores for the last three financial years	For Annual Turnover of Rs.6 Crores can be considered
		M/s Royal Enterprises	Request to consider Turnover Rs 10 crore., allow subdealer under authorised dealer to cater respective tender to all DMER medical college across Maharashtra.	OEM can appoint Only One Exclusive distributor
2	2.10 Tenderer must submit an Experience Certificate of ₹ 7 crore per year for the past three financial years, i.e. 2022-23, 2023-24, and 2024-25, duly signed by Maharashtra State Government institute purchasing authorities for satisfactory supply to their institute. (Pan India experience will be considered).	M/s Earthnine Naturals LLP	In previous tenders floated by the department, the experience criteria required submission of Government supply orders only. However, in the current tender, a requirement of ₹7 Crores has been specified. We request you to kindly align this with the earlier tender conditions with a maximum of ₹50 Lakhs per annum from Government institutions across India.	For Annual Turnover of Rs.6 Crores can be considered (Pan India experience will be considered)
		M/s Royal Enterprises	Request to consider Experience certificate of 2 crore for past three years and provide format of Experience certificate.	
		M/s Advanced Medical Solutions	We request to amend the clause as :Manufacturer should be allowed to participate directly in the tender and may execute supplies through multiple authorized distributors as this is bulk purchase and supplied to be made on timely basis to entire Maharashtra.	
		M/s Mediearth lifecare Pvt Ltd	The experience criteria be considered on a Pan India basis, rather than restricted to Maharashtra State Government institutions	
3	Regarding Single Bidder Authorization (in case of OEM participation)	M/s Meril Endo Surgery Pvt Ltd.	If an Original Equipment Manufacturer (OEM) participates directly in the tender, it shall authorize only one bidder to supply the quoted products on its behalf. Authorization of more than one bidder by the same OEM, or submission of bids through multiple authorized bidders in addition to direct participation, shall lead to rejection of all such bids.	Accepted

Sr. No.	Tender Clause and No	Name of the Bidder	Query / Clarification	Department Remark and Signature
4	Regarding Single Bid per Manufacturer (in case of distributor participation)	M/s Meril Endo Surgery Pvt Ltd.	An OEM may either participate directly or authorize only one distributor/bidder to submit a bid on its behalf. Submission of multiple bids by the same OEM, either directly or through more than one authorized distributor/bidder, shall result in rejection of all such bids.	Accepted
5	Regarding quantity purchase or it is Rate contract of two years	M/s Healthium Medtech Limited	Clarification required, whether it is quantity purchase or it is Rate contract of two years	Approximate required qty for Rate contract for two years
		M/s Johnson & Johnson Private Limited	Clarification required, whether it is quantity purchase or it is Rate contract of two years.	
6	Regarding revised manual of Gr dated 01.12.2016 is applicable for this tender or not	M/s Healthium Medtech Limited	Clarification required whether revised manual of Gr dated 01.12.2016 is applicable for this tender or not	Gr. Dated 01.12.2016 and all other government direction regarding tender and procurement will be applicable.
		M/s Johnson & Johnson Private Limited	Clarification required whether revised manual of Gr dated 01.12.2016 is applicable for this tender or not	
7	6.3 Envelope No. 2 (Price bid): <b>6.3 (D)</b> Financial bid of bidders may be opened in view of emergency purchase situation but after financial bid opening document of all bidders will be checked and if L1 bidder document will be inappropriate then rates of L2 ,L3, L4 bidder will be considered whose document will be okay provided bidder match L1 rates	M/s Healthium Medtech Limited	Price bid of eligible bidders only to be opened post completion of technical evaluation. And also Please consider the actual pack size of the product and place the orders in standard pack size, instead of single piece qty.	No Change
		M/s Johnson & Johnson Private Limited	AFTER OPEING THE PRICE BID if L1 Bidder is declared as non-responsive then L2 BIDDERS RATES TO BE CONSIDERED AS L1	
8	Regarding Rate Matching	M/s Advanced Medical Solutions	The L2 bidder be considered as L1, without requiring rate matching to the disqualified L1 bidder's price	No Change
		M/s Mediearth lifecare Pvt Ltd	The L2 bidder be considered as L1, without requiring rate matching to the disqualified L1 bidder's price	

Sr. No.	Tender Clause and No	Name of the Bidder	Query / Clarification	Department Remark and Signature
9	2.3 Manufacturer/Authorized dealer should produce Authority Letter and Only One Exclusive Authorized distributor is allowed	M/s Johnson & Johnson Private Limited	Requested to allow manufacturer to quote the tender by appointing area wise / region wise multiple Authorized distributors , so that supply can be made immediately on receipt of the order.	No Change
		M/s Advanced Medical Solutions	We request to amend the clause as :Manufacturer should be allowed to participate directly in the tender and may execute supplies through multiple authorized distributors as this is bulk purchase and supplied to be made on timely basis to entire Maharashtra.	
		M/s Mediearth lifecare Pvt Ltd	"Manufacturer should be allowed to participate directly in the tender and may execute supplies through multiple authorized distributors"	
10	2.4 Performance statement for last three years i.e.2022-23,2023-24 & 2024-25 supported by copies of purchase orders issued by any state	M/s Healthium Medtech Limited	it is requested to amend "Experience certificate of Rs.7 crore to any state government Avarage of the past three financial years, i.e. 2022-23,2023-24 & 2024-25 of the manufacturer, signed by manufacturer to be accepted."	Pan India experience from Government/ SemiGovnment will be considered.
		M/s Johnson & Johnson Private Limited	Any medical device products experience certificate certified by bidder/ manufacturer should be accepted. Experience certificate signed by manufacturer to be accepted.	
		M/s Advanced Medical Solutions	The experience criteria be considered on a Pan India basis, rather than restricted to Maharashtra State Government institutions. Also this certificate signed by Bidder /Manufacturer to be accepted	
		M/s Mediearth lifecare Pvt Ltd	The experience criteria be considered on a Pan India basis, rather than restricted to Maharashtra State Government institutions	
11	Request for EMD Relaxation/Exemption	M/s Aegis LifeSciences	we request exemption from EMD and tender fee for eligible bidders in accordance with Rule 170(i) of GFR 2017 or Reducing the EMD amount to a reasonable level (e.g., Rs.1,00,000/-)	No Change

  
 Director  
 Medical Education & Research  
 Mumbai 400001

**ANNEXURE-4 Part I**  
**(PRICE BID FOR THE CURRENT TENDER)**

*(To be submitted in Envelope No. 2 as PDF)*

Item Description	Unit	Quantity	Ex-factory cost per unit	GST applicable for Govt. Supply (In Rs.)	Other incidental charges (please specify) (In Rs.)	Total landed cost per unit (4+5+6) (In Rs.)	Total Cost Rs. (3 x7)
1	2	3	4	5	6	7	8

Total tender price (in words.....)

The price should be quoted only in Indian currency Note:

**In case of discrepancy between unit price and total price, the unit price shall prevail.** Only total landed cost per unit considered for rate comparison.

Signature of the Tenderer

Name

Designation

Business address

**To be uploaded in the form of Excel**

**ANNEXURE-4 PART II**

(Statement showing comparative prices offered by the tenderer in other tenders of the same product)

**ONLY FOR ADDITIONAL INFORMATION AS TO RATES OFFERED BY THE TENDERER IN VARIOUS OTHER TENDES.**

Please mention quoted rates of above item of different years

Sr. No.	Year	MRP per unit	Price as per DPCO /NPPA	Unit Price offered in other tenders/rate contracts					
				DHS (Govt. of Maharashtra)	DMER (Govt. of Maharashtra)	MCGM	ESIC	Other State Govt.(s)	Tenders of Central Govt.
1	2022-23								
2	2023-24								
3	2024-25								

Additional rows for information of other years can be inserted

**To be uploaded in Envelope no. 2 in PDF Format only.**

Signature

Seal