

Maharashtra Medical Goods Procurement Authority, Mumbai

E-253 Pre-Bid Query (Itemwise)

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
3	GOLDWIN MEDICARE LIMITED	3	Gloves General Examination	Disposable examination gloves made of soft latex, powder-free; highly elastic and tear-resistant; texturing of the finger area ensures optimum grip; excellent fit and high tactile sensitivity; With ASTM, ISO & C E approved	<p>1. Size of the Gloves to be clarified whether small/Medium/Large .</p> <p>2. It should be amended ISO OR CE approved.</p>	Size not mentioned and CE is nit applicable for DOMESTIC SALES . These products are reserved only for MSME Units, CE certificate should be relaxed for SSI Units as not exporting any products.	Size - small/Medium/Large .	<p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.</p>	139 R
								<p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.</p>	139
1	Adeshwar Meditex Ltd.	5	Gloves Surgical Rubber no 6	Surgical Rubber Gloves no. 6 Pair	The details specification is not mentioned, as the next sizes i.e. 6.5,7 are mentioned non sterile , it should be clarified whether it is non sterile or sterile ?	Rubber gloves 6. 'Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Cornstarch Powder, - ,ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)	Rubber gloves 6 Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
	Trusted Medical Equipments Industries	5	Gloves Surgical Rubber no 6	Surgical Rubber Gloves no. 6 Pair	<p>We are interested in participating in the tender process mentioned above. We require clarification regarding certain aspects of Annexure X – Compliance Sheet for Technical Proposal in the said tender.</p> <p>The technical specifications for the following item numbers are not completely and clearly specified / appear to be insufficient. Therefore, please clarify all the necessary technical criteria for the following items (such as ASTM/ISO/CE certifications, Sterile/Non-Sterile, Powdered/Powder-Free, etc.) so that we can submit a proper and accurate technical proposal.</p>		Rubber gloves 6 Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
3	GOLDWIN MEDICARE LIMITED	5	Gloves Surgical Rubber no 6	Surgical Rubber Gloves no. 6 Pair	If should amended: Rubber gloves 6.0 (Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, Conforms to IS/IS certification , Bio Absorbable Corn Starch Powder, ASTM - D .	There is no specification whether sterile or non sterile is mentioned .	Rubber gloves 6 Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R

**E-253 Pre-Bid Query (Itemwise)**

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
4	VERO GROUP	5	Gloves Surgical Rubber no 6	Surgical Rubber Gloves no. 6 Pair	<p>With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.</p> <p>The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements.</p> <p>We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.</p>	<p><b>With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.</b></p> <p><b>The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines.</b></p> <p><b>In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements.</b></p> <p><b>We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.</b></p>	Rubber gloves 6 Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष-दिनांक:२९.०८.२०२५.	139 R
4	VERO GROUP	6	Gloves Surgical Rubber no 6.5	Rubber gloves 6.5 Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D	<p>With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.</p> <p>The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements.</p> <p>We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.</p>	<p>With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.</p> <p>The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines.</p> <p>In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements.</p> <p>We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.</p>	Rubber gloves 6.5" Operation Gloves (non-sterilised) 6.5" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१२७/खरेदी कक्ष- दिनांक: १८.०६.२०२५.	127
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष-दिनांक:२९.०८.२०२५.	139 R

E-253 Pre-Bid Query (Itemwise)

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
4	VERO GROUP	7	Gloves Surgical Rubber no 7	Rubber gloves 7 Operation Gloves (non-sterilised) 6" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.	Rubber gloves 7" Operation Gloves (non-sterilised) 7" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३७/खरेदी कक्ष- दिनांक: १८.०६.२०२५.	127
					The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.		शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
1	Adeshwar Meditex Ltd.	8	Gloves Surgical Rubber no 8	Surgical Rubber Gloves no. 8 Pair	The details specification is not mentioned, as the earlier sizes i.e. 6.5,7 are mentioned non sterile , it should be clarified whether it is non sterile or sterile ?	Rubber gloves 8 Operation Gloves (non-sterilised)8 " pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Cornstarch Powder, <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)</b>	Rubber gloves 8" Operation Gloves (non-sterilised) 8" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
2	Trusted Medical Equipments Industries	8	Gloves Surgical Rubber no 8	Surgical Rubber Gloves no. 8 Pair	We are interested in participating in the tender process mentioned above. We require clarification regarding certain aspects of Annexure X – Compliance Sheet for Technical Proposal in the said tender. The technical specifications for the following item numbers are not completely and clearly specified / appear to be insufficient. Therefore, please clarify all the necessary technical criteria for the following items (such as ASTM/ISO/CE certifications, Sterile/Non-Sterile, Powdered/Powder-Free, etc.) so that we can submit a proper and accurate technical proposal.		Rubber gloves 8" Operation Gloves (non-sterilised) 8" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R

**E-253 Pre-Bid Query (Itemwise)**

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
3	GOLDWIN MEDICARE LIMITED	8	Gloves Surgical Rubber no 8	Surgical Rubber Gloves no. 8 Pair	If should amended: Rubber gloves 7 (Operation Gloves (non-sterilised) 7 " pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, Conforms to IS/IS certification , Bio Absorbable Corn Starch Powder, ASTM - D .	<b>No specification is mentioned.</b>	Rubber gloves 8" Operation Gloves (non-sterilised) 8" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
4	VERO GROUP	8	Gloves Surgical Rubber no 8	Surgical Rubber Gloves no. 8 Pair	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards. The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards. The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	Rubber gloves 8" Operation Gloves (non-sterilised) 8" pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Corn Starch Powder, ASTM - D Standard with NABL Test Report	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
1	Adeshwar Meditex Ltd.	9	Gloves Surgical Sterile no 6	Surgical Gloves Sterile no. 6 Pair	Specification should be clarified whether it is powdered or powder free ?	Sterile Disposable Surgical Gloves powder free, Size 6" , Conforms to ISI /ISI certification 13422 . <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO), Sterilized</b>	Sterile Disposable Surgical Gloves No 6" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139

## E-253 Pre-Bid Query (Itemwise)

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
2	Trusted Medical Equipments Industries	9	Gloves Surgical Sterile no 6	Surgical Gloves Sterile no. 6 Pair	We are interested in participating in the tender process mentioned above. We require clarification regarding certain aspects of Annexure X – Compliance Sheet for Technical Proposal in the said tender. The technical specifications for the following item numbers are not completely and clearly specified / appear to be insufficient. Therefore, please clarify all the necessary technical criteria for the following items (such as ASTM/ISO/CE certifications, Sterile/Non-Sterile, Powdered/Powder-Free, etc.) so that we can submit a proper and accurate technical proposal.		Sterile Disposable Surgical Gloves No 6" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139
3	GOLDWIN MEDICARE LIMITED	9	Gloves Surgical Sterile no 6	Surgical Gloves Sterile no. 6 Pair	<b>Sterile Disposal Surgical Gloves powder free ,Size 6" , Conforms to IS/IS certification.</b>	<b>Specification to be clarified</b>	Sterile Disposable Surgical Gloves No 6" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139
4	VERO GROUP	9	Gloves Surgical Sterile no 6	Surgical Gloves Sterile no. 6 Pair	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards. The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards. The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	Sterile Disposable Surgical Gloves No 6" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139

**E-253 Pre-Bid Query (Itemwise)**

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
1	Adeshwar Meditex Ltd.	10	Gloves Surgical Sterile no 7	Surgical Gloves Sterile no. 7 Pair	Specification should be clarified whether it is powdered or powder free?	Sterile Disposable Surgical Gloves powder free, Size 7" , Conforms to ISI/ ISI 13422 . <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO), Sterilized</b>	Sterile Disposable Surgical Gloves No 7" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139
2	Trusted Medical Equipments Industries	10	Gloves Surgical Sterile no 7	Surgical Gloves Sterile no. 7 Pair	We are interested in participating in the tender process mentioned above. We require clarification regarding certain aspects of Annexure X – Compliance Sheet for Technical Proposal in the said tender. The technical specifications for the following item numbers are not completely and clearly specified / appear to be insufficient. Therefore, please clarify all the necessary technical criteria for the following items (such as ASTM/ISO/CE certifications, Sterile/Non-Sterile, Powdered/Powder-Free, etc.) so that we can submit a proper and accurate technical proposal.		Sterile Disposable Surgical Gloves No 7" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139
3	GOLDWIN MEDICARE LIMITED	10	Gloves Surgical Sterile no 7	Surgical Gloves Sterile no. 7 Pair	<b>Sterile Disposal Surgical Gloves powder free , Size 7",Conforms to IS/IS certification .</b>	<b>Specification to be clarified</b>	Sterile Disposable Surgical Gloves No 7" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९/खरेदी कक्ष- दिनांक: १२.०६.२०२५.	139

## E-253 Pre-Bid Query (Itemwise)

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
4	VERO GROUP	10	Gloves Surgical Sterile no 7	Surgical Gloves Sterile no. 7 Pair	<p>With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.</p> <p>The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements.</p> <p>We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.</p>	<p>With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards.</p> <p>The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements.</p> <p>We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.</p>	<p>Sterile Disposable Surgical Gloves No 7" Powder free gloves, polymer coated inside with powder content &lt;2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark</p>	<p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.</p>	139 R
1	Adeshwar Meditex Ltd.	11	Gloves Surgical Sterile no 8	Surgical Gloves Sterile no. 8 Pair	<p>Specification should be clarified whether it is powdered or powder free?</p>	<p>Sterile Disposable Surgical Gloves powder free, <b>Size 8"</b> , Conforms to ISI/ISI certification 13422 , . <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO), Sterilized</b></p>	<p>Sterile Disposable Surgical Gloves No 8" Powder free gloves, polymer coated inside with powder content &lt;2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark</p>	<p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.</p>	139 R
2	Trusted Medical Equipments Industries	11	Gloves Surgical Sterile no 8	Surgical Gloves Sterile no. 8 Pair	<p>We are interested in participating in the tender process mentioned above. We require clarification regarding certain aspects of Annexure X – Compliance Sheet for Technical Proposal in the said tender. The technical specifications for the following item numbers are not completely and clearly specified / appear to be insufficient. Therefore, please clarify all the necessary technical criteria for the following items (such as ASTM/ISO/CE certifications, Sterile/Non-Sterile, Powdered/Powder-Free, etc.) so that we can submit a proper and accurate technical proposal.</p>		<p>Sterile Disposable Surgical Gloves No 8" Powder free gloves, polymer coated inside with powder content &lt;2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark</p>	<p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.</p>	139 R

**E-253 Pre-Bid Query (Itemwise)**






Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
3	GOLDWIN MEDICARE LIMITED	11	Gloves Surgical Sterile no 8	Surgical Gloves Sterile no. 8 Pair	<b>Sterile Disposal Surgical Gloves powder free , Size 8"Conforms to IS/IS certification .</b>	<b>Specification to be clarified</b>	Sterile Disposable Surgical Gloves No 8" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R
4	VERO GROUP	11	Gloves Surgical Sterile no 8	Surgical Gloves Sterile no. 8 Pair	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards. The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	With reference to the above-mentioned tender, we would like to bring to your kind notice that the BIS standard IS 4148, which was earlier applicable for Surgical Gloves, has now been withdrawn/expired by the Bureau of Indian Standards. The revised and currently applicable BIS standard for Surgical Gloves is IS 13422, as per the latest BIS notifications and guidelines. In view of the above, we kindly request you to update/replace the BIS standard from IS 4148 to IS 13422 in the tender documents/records for Surgical Gloves, in order to ensure compliance with the current BIS regulatory requirements. We assure you of our full cooperation and are ready to submit any additional documents or clarifications, if required.	Sterile Disposable Surgical Gloves No 8" Powder free gloves, polymer coated inside with powder content <2mg as per USFDA. 100% electronically tested. CE 2460certification (Notified body European Standard), EN 455-2 OR ISI Mark	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R

**E-253 Pre-Bid Query (Itemwise)**

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Ammendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
1	Adeshwar Meditex Ltd	12	Sanitary Napkin	<p>Sanitary Napkin consists of an outer covering provided with sufficient number of channels for leak protection and an absorbent filter material with an adhesive back strip. B. Description 1.covering-The covering of the absorbent filler shall be made of good quality perforated polyurethane film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. This shall be made of a product that is non-allergenic and bio-compatible. The sanitary napkins shall have non-absorbent barrier on one side which shall have an identifying mark. Indicating clearly the side of barrier. 2. Absorbent Filler- The filler material shall consist of cellulose pulp (either based on wood or paper or poly-acrylate gels). This shall be free from lumps, oil spots, dirt or foreign material etc. 3. Back Strip- A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4. Absorbency- the sanitary napkin should be able to absorb not less than 30ml of normal saline (I.P.) @ 15ml per minute. 5. Size -The size of absorbent section of the Sanitary Napkin shall be as follows: Pad Length 235+/- 10mm, Core length 210+10mm, overall width 155+ 5mm, core width 65+ 5mm, Pad Thickness 7-9mm The thickness shall be measured by stacking 10 complete pads and measuring the stack height</p>	<p>Pack of how many pads required is not clarified i.e. whether pack of 5/6/ 8 , which pack is required .It will help us to quote the rates accordingly</p>	<p>Sanitary Napkin consists of an outer covering provided with sufficient number of channels for leak protection and an absorbent filter material with an adhesive back strip. Description covering-The covering of the absorbent filler shall be made of good quality perforated polyurethane film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. This shall be made of a product that is non-allergenic and bio-compatible. The sanitary napkins shall have non-absorbent barrier on one side which shall have an identifying mark. Indicating clearly the side of barrier. 2. Absorbent Filler- The filler material shall consist of cellulose pulp (either based on wood or paper or poly-acrylate gels). This shall be free from lumps, oil spots, dirt or foreign material etc. 3. Back Strip- A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4. Absorbency- the sanitary napkin should be able to absorb not less than 30ml of normal saline (I.P.) @ 15ml per minute. 5. Size -The size of absorbent section of the Sanitary Napkin shall be as follows: Pad Length 235+/- 10mm, Core length 210+10mm, overall width 155+ 5mm, core width 65+ 5mm, Pad Thickness 7-9mm The thickness shall be measured by stacking 10 complete pads and measuring the stack height, <b>Pack of 6 pads . . ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)</b></p>	<p>Description :  <b>1. Covering</b> – The covering of the absorbent filler shall be made of good quality perforated Polyurethane film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. This shall be made of a product that is non-allergenic and bio – compatible. The sanitary napkins shall have a non-absorbent barrier on one side which shall have an identifying mark indicating clearly the side of the barrier.  <b>2. Absorbent Filler</b> – The filler material shall consist of cellulose pulp (either based on wood or paper or poly-acrylate gels). This shall be free from lumps, oil spots, dirt or foreign material etc.  <b>3. Back Strip</b> – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material.  <b>4. Absorbency</b> – The sanitary napkin should be able to absorb not less than 30 ml of normal saline (I.P.) @ 15 ml/minute.  <b>5. Size</b> – The size of absorbent section of the Sanitary Napkin shall be as follows: Pad Length Core Length Overall Width Core Width Pad Thickness 235+/- 10mm 210+/- 10mm 155+/- 5mm 65+/- 5mm 7-9mm                      The thickness shall be measured by stacking 10 complete pads and measuring the stack height. The average thickness for the 10 pads shall be used as the pad thickness.  <b>6. Weight</b> – The weight of one full sanitary napkin shall not be more than 7.5 – 9 grams.  <b>7. pH Value</b> – The sanitary napkin shall be free from acids and alkali and the pH of the absorbent material shall be 6 – 8.5 when tested by the method given in IS : 1390 – 1961 (methods for determination of pH value of aqueous extracts of textile materials.)  <b>PACK OF 6 No</b></p>	<p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१२७/खरेदी कक्ष- दिनांक: १८.०६.२०२५.</p> <p>शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक: २९.०८.२०२५.</p>	<p>127</p> <p>139 R</p>

**E-253 Pre-Bid Query (Itemwise)**

Sr. No.	Name of the Bidder	Item Sr. No. (As per Appendix A)	Name of the Item	Specifications	Query / Clarification	Amendment Request	Department Remark	Concern Department	Prama no
1	2	3	4	5	6	7	8	9	10
3	GOLDWIN MEDICARE LIMITED	12	Sanitary Napkin	Sanitary Napkin consists of an outer covering provided with sufficient number of channels for leak protection and an absorbent filter material with an adhesive back strip. B. Description 1.covering-The covering of the absorbent filler shall be made of good quality perforated polyurethane film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements.This shall be made of a product that is non-allergenic and bio-compatible.The sanitary napkins shall have non-absorbent barrier on one side which shall have an identifying mark.Indicating clearly the side of barrier. 2.Absorbent Filler- The filler material shall consist of cellulose pulp(either based on wood or paper or poly-acrylate gels).This shall be free from lumps,oil spots,dirt or foreign material etc. 3.Back Strip-A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4.Absorbency- the sanitary napkin should be able to absorb not less than 30ml of normal saline( I.P.)@15ml per minute. 5.Size -The size of absorbent section of the Sanitary Napkin shall be as follows: Pad Length 235+- 10mm,Core length 210+10mm,overall width 155+ 5mm,core width 65+ 5mm, Pad Thickness 7-9mm The thickness shall be measured by stacking 10 complete pads and measuring the stack height	Packing is not mentioned it should be clarified packet of 6 pads or 8 pads or 10 pads.	Normally the packet of six pads are procured .	Description : 1.Covering – The covering of the absorbent filler shall be made of good quality perforated Polyurethane film sleeve which has sufficient porosity to permit the assembled napkin to meet the absorbency requirements. This shall be made of a product that is non-allergenic and bio – compatible. The sanitary napkins shall have a non-absorbent barrier on one side which shall have an identifying mark indicating clearly the side of the barrier. 2. Absorbent Filler – The filler material shall consist of cellulose pulp (either based on wood or paper or poly-acrylate gels). This shall be free from lumps, oil spots, dirt or foreign material etc. 3. Back Strip – A back strip for sticking the sanitary napkin onto the underwear should be there using good quality adhesive material. 4. Absorbency – The sanitary napkin should be able to absorb not less than 30 ml of normal saline (I.P.) @ 15 ml/minute. 5. Size – The size of absorbent section of the Sanitary Napkin shall be as follows: Pad Length Core Length Overall Width Core Width Pad Thickness 235+ _10mm 210+_10mm 155+_5mm 65+_5mm 7-9mm The thickness shall be measured by stacking 10 complete pads and measuring the stack height. The average thickness for the 10 pads shall be used as the pad thickness. 6. Weight – The weight of one full sanitary napkin shall not be more than 7.5 – 9 grams. 7. pH Value – The sanitary napkin shall be free from acids and alkali and the pH of the absorbent material shall be 6 – 8.5 when tested by the method given in IS : 1390 – 1961 (methods for determination of pH value of aqueous extracts of textile materials.) PACK OF 6 No	शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१२७/खरेदी कक्ष- दिनांक: १८.०६.२०२५.	127
								शासन निर्णय क्रमांक: प्रशामा-१२२५/प्र.क्र.१३९ (सुधारित)/खरेदी कक्ष- दिनांक:२९.०८.२०२५.	139 R

				
<b>Pharmacist</b>	<b>Senior. Pharmacy Officer</b>	<b>Technical Officer (Technical)</b>	<b>Chief Acc &amp; Finance Officer</b>	<b>General Manager (Technical)</b>
Miss. Ojasvi Devlekar	Mr. Dnyaneshwar Visawe	Dr. D. B. Khelkar	Mr. Bhagwat Murkute	Dr. Sanjeevkumar Jadhav
MMGPA, Mumbai	MMGPA, Mumbai	MMGPA, Mumbai	MMGPA, Mumbai	MMGPA, Mumbai

## Maharashtra Medical Goods Procurement Authority, Mumbai

### E-253 Pre-Bid Query (General Queries)

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
1	Adeshwar Meditex Ltd	11	2. Eligibility criteria:	<b>8) Average Annual Turnover</b> Average Annual Turnover (in last three financial years (2022-23,2023-24, 2024-25) (as attached Annexure-6)	Format enclosed in the tender form is Annexure VI contract form, and for Turnover annexure is mentioned is Annexure IV For Turnover instead of as mentioned Annexure VI to be amended as Annexure IV	<b>It is typographical error to be corrected</b>	Read as Annexure IV - Average Annual turnover
		12		<b>13) Certification</b> 1. Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9)/ ISO 13485 (QMS) Quality management System issued by Notified Body.	Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5 & ISO 13485 (QMS) Quality management System issued by Notified Body both are mandatory. i.e. instead of OR ,TO BE AMENDED AS "AND"	Actually as per the Medical device rules 2027, for manufacturing of Medical devices Both the Certificate I.e (MD- 5 & ISO 13485 (QMS) Quality management System issued by Notified Body both are mandatory.	Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9/) and ISO 13485 (QMS) Quality management System issued by Notified Body.
		51	Appendix A- Schedule of requirement,	<b>Packaging Turnover related query</b>	<b>Appendix A- Schedule of requirement, Packaging Turnover related query :-</b> Item No.12, Sanitary NAPKIN , Packaging Mentioned per piece, , it should be clarified whether required quantity is in pack of individual pieces or pack of 6 pads ? .	Item No.12, Sanitary NAPKIN , Packaging Mentioned per piece, and turnover mentioned 1508486250.00 ( assuming 50% Of the required value) i.e. per piece rate will be 12 per piece , Actually market rate per piece is note more than maximum Rs. 4.00 per piece. By increasing the value per piece, turnover also increased . Therefore it should be clarified whether the required quantity is per piece or pack of 5 or 6 pads ?	Quantity floated is per piece
		51	Appendix A- Schedule of requirement,	<b>Packaging Turnover related query</b>	<b>Item No. 13,14 &amp;15</b> – Pop , packaging mentioned per roll and item no 15 , mentioned packaging pack of 10 rolls, it should be clarified whether quantity mentioned is of per piece or pack of 10 rolls .	Assuming 50% Of the required value) , the turnover mentioned is higher , i.e. for the item no. 13, per roll rate will be 123.00 which is in higher side , and for item no. 14, per roll will be 172.00 and for item no. 15, per roll rate will be 940.00, HENCE IT IS TO BE CLARIFIED THE QUANTITY MENTIONED ARE PACK OF PER ROLL OR PACK OF SINGLE PIECE.	1) Tendre item no 13, 14 Packing is pack of 10 Rolls 2) Quantity floated is per roll in unit

**E-253 Pre-Bid Query (General Queries)**

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
1	Adeshwar Meditex Ltd	51	Appendix A-Schedule of requirement,	Absorbent Cotton Wool IP 500 mg	It should be CLARIFIED WHETHER 50MG OR 500GM ?	Absorbent Cotton Wool IP/BP/USP pack of 500 gm , Sterile, Approved as per Medical Device Rules 2017 , MD-5 Licence along with ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO) Size 500gm Roll , Sterile	Absorbent Cotton Wool IP/BP/USP pack of 500 gm , Sterile, Approved as per Medical Device Rules 2017 , MD-5 Licence along with ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO) Size 500gm Roll , Sterile
		12	2. Eligibility criteria:	<b>13) Certification</b> 1. Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9)/ ISO 13485 (QMS) Quality management System issued by Notified Body.	It should be Schedule F- II as per Drug and Cosmetic Act 1940 & ISO 13485 ; 2016 along with MD-5 licence is mandatory for Medical devices .	Absorbent Surgical Gauze Schedule F II as per drugs and cosmetic act 1940 & as approved by Medical Device Rules 2017 AS PER MD -5 Licence along with ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO) (50cmx18 Mtr) , <b>Sterilized</b>	Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9/) and ISO 13485 (QMS) Quality management System issued by Notified Body.
					CE Is applicable only for EXPORT PURPOSE, therefore it should be amended instead of CE to <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)</b>	Disposable examination gloves made of soft latex, powder-free; highly elastic and tear-resistant; texturing of the finger area ensures optimum grip; excellent fit and high tactile sensitivity; <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO), Sterilized.</b>	ASTM, ISO & FDA/CE 93/42/EEC
CE Is applicable only for EXPORT PURPOSE, therefore it should be amended instead of CE to <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)</b>	Disposable examination gloves made of soft latex, powder-free, highly elastic and tear-resistant, texturing of the finger area, ensures optimum grip, excellent fit and high tactile sensitivity, <b>ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO), Sterilized</b>	ASTM, ISO & FDA/CE 93/42/EEC					

## E-253 Pre-Bid Query (General Queries)

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
1	Adeshwar Meditex Ltd			Gloves Surgical Rubber no 6.5	Typographical error to be corrected as size mentioned both i.e. 6 & 6.5 , it should be clarified .	Rubber gloves 6.5 Operation Gloves (non-sterilised) 6.5 " pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Cornstarch Powder, ,ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)	Gloves surgical non sterilised sixe ;- 6.5
				Gloves Surgical Rubber no 7	Typographical error to be corrected as size mentioned both i.e. 7 & 6 , it should be clarified .	Rubber gloves 7 Operation Gloves (non-sterilised)7 " pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, IS certification with CE certification (Notified Body), Bio Absorbable Cornstarch Powder, , ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)	Gloves surgical non sterilised sixe ;- 7
				Plaster of Paris Bandage (B.P) Readymade Size 10 cm x 2.7 Mtrs	To be amended the Size 15 cm x 2.7 mtr,	Plaster of Paris Bandage (Readymade) Size 10 cm x 2.7 Mtrs. ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)	No change
				Plaster of Paris Bandages (Ready made) Size 15 cm x 2.7 m	To be amended the Size 15 cm x 2.7 mtr,	Plaster of Paris Bandages (Ready-made) Size 15 cm x 2.7 mtr . ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)	No change

**E-253 Pre-Bid Query (General Queries)**

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
1	Adeshwar Meditex Ltd			Plaster of Paris Bandages Size 20 cm x 2.7 m Roll	To be amended the Size 20 cm x 2.7 mtr	Plaster of Paris Bandages (Ready-made) Size <b>20 cm x 2.7 mtr . ISO 13485: 2016 certified (Certificate issued by Notified body approved by CDSCO)</b>	No change
2	M/s. Worldwide Surgical Cotton	11	<b>2. Eligibility criteria:</b>	<b>8) Average Annual Turnover</b> Average Annual Turnover (in last three financial years (2022-23,2023-24, 2024-25) (as attached Annexure-6)	This tender includes a total of 14 products (items) with a total estimated cost of ₹26,18,86,350/-. However, the tender specifies a separate product-wise annual turnover condition for each product (e.g., ₹1,84,39,750/-, ₹2,22,27,708/-, etc.). Our organization is a Micro Enterprise and is validly registered under the MSME Development Act, 2006 with a Udyam registration. According to Government Resolution No. MSME-2016/C.R.227/Industry-7, dated 01.12.2016, of the Industries, Energy and Labour Department, Government of Maharashtra, there are clear instructions to relax excessive financial conditions such as high annual turnover and experience requirements as much as possible, with a view to increasing the participation of Micro and Small Enterprises in the state government's procurement process. In light of the above, the imposition of a high product-wise turnover condition for each product separately is actually limiting the participation of Micro MSME manufacturers, and this condition is inconsistent with the objective of the above-mentioned government policy.	Therefore, please clarify whether the product-wise turnover condition in this tender will be relaxed for Micro MSME manufacturers, and whether a cumulative annual turnover of at least ₹2,00,00,000/- (Rupees Two Crore) will be accepted? Alternatively, please clarify whether this condition will be applied in a concessional manner for Micro MSMEs. Your guidance and clarification are requested.	No change
3	Laxmi Industries	12	<b>2. Eligibility criteria:</b>	<b>8) Average Annual Turnover</b> Average Annual Turnover (in last three financial years (2022-23,2023-24, 2024-25) (as attached Annexure-6)	As per Clause No.4.4.1.2 and 5.14 of Maharashtra Govt. G.R.Dt.01/12/2016 it is stated that 'no such clauses should not be mentioned in the terms and conditions of the tender which hamper competition and restrict eligibility of bidders' so keeping in mind that these items are now reserved to be purchased from Micro SSI units of the state we request you to reduce the Turnover criteria for MSME units. This was also conveyed to DHS, mumbai by The Directorate of Industries vide their Letter No. cspo/mssdlu/a-1408 dt.5/8/2016 and later HBPCL also kept the turnover for these items at Rs. One Crore in their tenders we once again request you to grant exemption for the same.		No change

## E-253 Pre-Bid Query (General Queries)

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Ammendment Request	Department Remark
		51	Appendix-A	(schedule of requirements)	As per Appendix-A (schedule of requirements) wherein the proposed quantities and required turnover are given; in case of Plaster of Paris Bandage, Item Nos.13,14 &15, the required turn over multiplied by two and divided by the required quantity gives exhorbent individual per piece rates of Rs.123/-,Rs.343/- & Rs.940, is it a typo error in printed Require Turnover, which is needed to check, further the Packing demonstrated for item No.15 only is pack of 10 Rolls, whereas for item Nos 13 & 14 says single Roll, please clarify.		Item no 13, 14 pack of 10 rolls rest no change
3	Laxmi Industries				<p>As per Govt. Resolution by Industries, Energy &amp; Labour Department, Maharashtra State, dated 1.12.2016 - Entities who are registered under Micro, Small and Medium Enterprises development act 2006 are exempted from paying Tender Form Fees and Earnest Money Deposits.</p> <p>As per the 5.14 Reservations and facilities for products manufactured from Micro, Small &amp; Medium Enterprises.</p> <p>IT is MENTIONED IN PARA :of clause no. 5.14. "Accordingly, while procuring 100% procurement of reserved items from Micro &amp; Small-Scale Industries, 20% of purchases should be reserved for the manufacturers belonging to scheduled caste &amp; scheduled tribes. Maharashtra State Small Industries Development Corporation can also participate in the said tender proceedings."</p> <p>In reference to the above we request to have following clarification:</p> <ol style="list-style-type: none"> <li>1. What documents must be uploaded in the tender for claiming the Privileges to manufacturers belonging to scheduled caste &amp; scheduled tribes., In case of proprietary MSE wherein the proprietor(s) are SC /ST.</li> <li>2. Any relaxation in Turnover clause &amp; Past performance clause is applicable to SC/ST firm.</li> <li>3. Is it mandatory to submit Caste Certificate and issue of such caste certificates validity verified by concerned issuing authority/ Scrutiny Committee about the genuineness on the time of submission of tender for availing such BENEFITS? Or Submission of Caste Validity is a Must.</li> </ol>		<ol style="list-style-type: none"> <li>1) SC/ST caste certificate with Caste validity certificate.</li> <li>2) No change</li> <li>3) For availing benefits of SC/ST submit catse certificate along with Caste validity certificate.</li> </ol>

**E-253 Pre-Bid Query (General Queries)**

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
3	Laxmi Industries	23	<b>29. Payment:</b>	<p>Payment against supply order issued under this bid will be made by Maharashtra Medical Goods Procurement Authority, Mumbai.                      100% Payment shall be made upon submission of following documents:                      (i) 2 copies of supplier's invoice.                      (ii) Receipt and acceptance certificates issued by the consignees.                      (iii) Batch wise In house Lab Report                      (iv) Payments towards the supply of Drugs will be made strictly as per the rules of MMGPA, Mumbai. The payment will be made through RTGS/ NEFT. The bidder shall furnish the relevant details in original (Annexure 8) to make the payment through RTGS/NEFT and the change of Bank Account during the validity of the bid will not be entertained normally.                      The purchaser shall have every right to deduct the pending dues on account of loss, compensation, or any remedial action in monetary terms from the said payment. The supplier shall not agitate the said issue in future.</p>	<p>As per Clause No. 30(4)- Payments : It is mentioned that "Payments towards the supply of goods will be made strictly as per the rules of MMGPA, Mumbai. " But the rules of Authority is not disclosed. Hence it is requested to publish the time frame for Release of Payment as, As per the MSME Act, Section 43B(h) Time Limit: Business enterprises are required to pay MSMEs within 45 days, as per section 15 of the MSMED Act 2006, depending on the presence of a written agreement. In case there is no written agreement, payment should be made within 15 days. In case there is a written agreement, payment shall be made as per the agreed-upon time line, not exceeding 45 days.</p>		Payment will be done as per clause no 29 of the RFP document.
					<p>We request you to provide grace period for submissions of discrepancies in Technical documents, as it is not mention in the T&amp;C of the Tender referred above Tender.</p>		No change
4	GOLDWIN MEDICARE LIMITED	11	<b>2. Eligibility criteria:</b>	<b>8) Average Annual Turnover</b> Average Annual Turnover (in last three financial years (2022-23,2023-24, 2024-25) (as attached Annexure-6)	<b>Format enclosed in the tender form is Annexure VI contract form, and for Turnover annexure is mentioned is Annexure IV For Turnover instead of as mentioned Annexure VI to be amended as Annexure IV</b>	<b>It is typographical error to be corrected</b>	Read as Annexure IV - Average Annual turnover
		13		<b>13) Certification</b> 1. Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9)/ ISO 13485 (QMS) Quality management System issued by Notified Body.	<p>Actually as per the Medical device rules 2027, for manufacturing of Medical devices Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9) &amp; ISO 13485 (QMS) Quality management System issued by Notified Body both are mandatory.                      It should be clarified whether any one of the document will be accepted or . Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9) &amp; ISO 13485 (QMS) Quality management System issued by Notified Body REQUIRED .</p>	<b>It should be amended to &amp; instead of OR , i.e. Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9) &amp; ISO 13485 (QMS) Quality management System issued by Notified Body.</b>	Certificate issued by appropriate Licensing Authority as per Medical Device Rules 2017(MD-5/MD-9) and ISO 13485 (QMS) Quality management System issued by Notified Body.
				Gloves Surgical Rubber no 6.5	<p>If should amended:                      Rubber gloves 6.5 (Operation Gloves (non-sterilised) 6.5 " pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, Conforms to IS/IS certification , Bio Absorbable Corn Starch Powder, ASTM - D .</p>	<b>There is typographically error as mentioned in both 6.5 and 6 both sizes. to be corrected</b>	Gloves surgical non sterilised size ; - 6.5

## E-253 Pre-Bid Query (General Queries)

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
4	GOLDWIN MEDICARE LIMITED			Gloves Surgical Rubber no 7	If should amended: Rubber gloves 7 (Operation Gloves (non-sterilised) 7 " pre-powdered, BIS 4148 Specifications gloves, surgical rubber made from Natural Latex. 100% electronically tested, Conforms to IS/IS certification , Bio Absorbable Corn Starch Powder, ASTM - D .	<b>There is typographically error as mentioned in both 7 and 6 both sizes. to be corrected</b>	Gloves surgical non sterilised sixe ;- 7
5	Natira Industries Pvt. Ltd.	51	Appendix-A	(schedule of requirements)	Government Resolution no. Bha.kha.sa-2014/ Pra.Kra: 82/Sec□on-III/Industry-4, DATED 01.12.2016 as per the clause no. 3.2.4 - Limited Tenders / Restricted Tenders- Point No. 4. The tenderer should not be awarded the contract worth more than its last financial year turn over. In case the value of procurement exceeds the last financial year turnover of the wining tenderer, the addi□onal order value should be awarded to the next eligible tenderer provided he matches the L1 price, Our turnover is lesser than the required turnover, whether we are eligible to par□cipate in the tender or not ?, if yes, what value of the contract will be awarded to us , whether up to the value of our last year turnover or the value of the total contract ?. If the clarifica□on is given can par□cipate in the tender accordingly.		No change
6	M/s. SKY INDUSTRIES	11	2. Eligibility criteria:	8) Average Annual Turnover Average Annual Turnover (in last three financial years (2022-23,2023-24, 2024-25) (as attached Annexure-6)	A) As per Clause No.4.4.1.2 and 5.14 of Maharashtra Govt. G.R.Dt.01/12/2016 it is stated that 'no such clauses should not be mentioned in the terms and conditions of the tender which hamper competition and restrict eligibility of bidders' so keeping in mind that these items are now reserved to be purchased from Micro SSI units of the state we request you to reduce the Turnover criteria for MSME units. This was also conveyed to DHS, Mumbai by The Directorate of Industries vide their Letter No. CSPO/MSSDLU/A-1408 Dt.05/08/2016 and the later HBPCL also kept the turnover for these items at Rs. One Crore in their tenders we once again request you to grant exemption for the same.  The T&C says Annexure -IV: Average Annual Turnover of the last three years along with item wise turnover will be taken as qualifying criteria along with +ve Net worth of the bidder. This clause of item wise turn-over is unwarranted and request you to delete it as as per the Gr Dt. 23/07/2021, the products are reserved for upliftment of MICRO & SMALL enterprises, so in order to accommodate all of them we request you to Company Annual Turn Over instead of item wise turnover as qualifying criteria and also relax the condition of+ve net worth for them, as most of the MSME are still running under loss due to COVID stress and will be restricted from participating and the Govt's Motto behind Reserving these items for purchase from them will not be served.		No change  No change

**E-253 Pre-Bid Query (General Queries)**

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
6	M/s. SKY INDUSTRIES	12	<b>2. Eligibility criteria:</b>	<b>10) Past Work Experience</b> The bidder must submit particulars of quantity of the past supplies made as per Annex 3. Out of this at least 25 % quantity for “Similar item as specified in the Technical Specification and in the Schedule of Requirements” must have been supplied in any of the last 3 (Three) Financial years i.e. 2022-23, 2023-24,2024-25	As per T&C for 'Past Work Experience' the bidder must submit production sale data of at least 25% in any of the past three years along with order copies, consignee details, consignee order nos., consignee item wise quantity supplied along with its value in 'Annexure-3", we request you to amend this condition of supply as this condition has been relaxed to 10% for Micro SSI units of the state vide G.R.No.HBP-2018/Pra.Kra.77/18/Medicine-1 dt.30/08/2018 by Medical Education and Drugs Department also the format demanded in Annexure-III is against the confidentiality of the tender, as by enclosing the order copies our consignee rates will be disclosed to our contenders. Secondly by asking to upload the purchase order etc. is not possible as the "GB" (space for uploading of the documents) allowed in the portal is limited and if further information required, documentary evidence in support of its credentials such as agreement copy/ work order / Letter of Award, work completion certificate, customer satisfaction certificates can be asked while technical scrutiny if required. Further CA will take their own time for verification and certification of such certificates, Therefore only Annexure -III certified by CA with UDIN Number with production and sale data only will serve the purpose as was done till date Hence we request you to Amend the format of Annexure-III. As per sample given below		No change
		21		<b>23) Sample Submission:</b>	As per clause No.23 of the T&C of the Tender referred above, the successful Bidder have to submit samples to MMGP after successful bid submission. The sample submission for testing and evaluation was abolished by DMER, DHS and HBPC LTD for surgical dressings as surgical dressings as they are made to pharmacopeial specifications and the technical committees don't have the infrastructure to test them and this condition may lead to Malpractices and Degrade the Authority so hereby request you to delete this condition for surgical dressings as Test Reports are submitted at the time of supplies.		No change
		28	2.1	<b>10) Past Work Experience</b>	The T&C of the Tender says Experience of bidder should be supported by CA Certificate with UDIN, but submission of Annexure-III suffices the need and request you to accept the same and avoid submission of bulky documents.		No change

Sr.No.	Year	Name and full address of the purchaser	Name of the Product	Batch No.	Quantity	
					Manufactured	Sold

## E-253 Pre-Bid Query (General Queries)






Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
6	M/s. SKY INDUSTRIES	23	29 (ii)	<b>29) Payments</b> (ii) Receipt and acceptance certificates issued by the consignees.	The Clause 29 of the T&C of the Tender referred above, says 100% Payment will be made upon Receipt of 'Receipt and Acceptance Certificates issued by the Consignees', we request you to kindly omit this sub clause as getting acceptance certificates is not in the domain of the supplier and it's the duty of the respective consignee to provide it in prescribed time and Micro SSI units have no machinery to collect it. So, we once again request you to grant exemption for the same.		No change
		23	29. (iv)	<b>29) Payments</b> (iv) Payments towards the supply of Drugs will be made strictly as per the rules of MMGPA, Mumbai. The payment will be made through RTGS/ NEFT. The bidder shall furnish the relevant details in original (Annexure 8) to make the payment through RTGS/NEFT and the change of Bank Account during the validity of the bid will not be entertained normally.	As per Clause No. 30(iv) - Payments: It is mentioned that "Payments towards the supply of goods will be made strictly as per the rules of MMGPA, Mumbai." But the rules of Authority is not disclosed. Hence it is requested to publish the time frame for Release of Payment as, As per the MSME Act, Section 43B (h) Time Limit: Business enterprises are required to pay MSMEs within 45 days, as per section 15 of the MSMED Act 2006, depending on the presence of a written agreement. In case there is no written agreement, payment should be made within 15days. In case there is a written agreement, payment shall be made as per the agreed-upon time line, not exceeding 45 days.		As per clause no 29
					We request you to provide grace period for submissions of discrepancies in Technical documents, as it is not mention in the T&C of the Tender referred above Tender.		No change
					We request you, as per directions given by Industries Department of Maharashtra to your Department on Dt. 04/10/2018 to grant 50% Purchase orders to L1 and 50% Purchase orders to eligible bidders in tenders. Kindly consider our Suggestions and Request for Amendments in Pre-Bid Queries compliance which will go a long way in ensuring the survival of Micro and Small SSI units of Maharashtra.		No change
7	Herambh Healthcare				We request you to kindly allow a supply period of 120-150 days instead of 45 days for the supply of 2,50,50,000 sanitary pads.  The currently stipulated period of 45 days is insufficient for such a large quantity. An extended timeline will enable us to manage the supply in a timely, disciplined manner while ensuring proper quality assurance at every stage.  We assure you of our full commitment to maintaining quality standards and completing the supply within the revised timeframe.		No change

**E-253 Pre-Bid Query (General Queries)**

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Ammendment Request	Department Remark
8	Sewagram Surgical Products				<p>१. आपले औषधी खरेदी घेण शासन निर्णय दि.२८/१२/२०१६ च्या परिपत्रक ४५.१.२ अन्वये निविदेसाठी पूर्वअर्हतेचे निकषामध्ये खरेदी कार्यालय अशा कोणतीही अट घालू शकत नाही, मात्र सदर अटीअंतर्गत खरेदी केलेल्या निविदाकारास समाधानकारक राहिल आणि जी निविदाकारामध्ये भेदभाव करणारी असेल अशी तरतूद असून सुद्धा पूर्वअर्हतेच्या निकषानुसार निविदा E-94 मध्ये वार्षिक उलाढाल व उत्पादन क्षमता मात्रा शिथिल करण्यात आलेली नाही. तरी सदर संदर्भात शासन निर्णय क्र.२०२०/उद्योग-६ दि.२३ जुलै २०२२ च्या संदर्भात आलेल्या अंमलबजावणीकरण करणेकरीता शासन निर्णय दि.२८/१२/२०१६ परिपत्रक ४५.१.२ अन्वये उत्पादन क्षमता शिथिल करण्याबाबत शासन निर्णयांची अंमलबजावणी करणेकरीता निविदा E-253 मध्ये वार्षिक उलाढाल व उत्पादन क्षमता मात्रा शिथिल करून L1 च्या दर मंजूर करण्याची इतर निविदाकारांस संधी देण्यात यावी.</p> <p>२. आम्ही उत्पादन मेडिकल डिव्हिजनमध्ये येत असल्यामुळे निविदेच्या अटी व शर्ती क्र.२.१.२ अन्वये उत्पादनांचे सॅम्पल सेंटरलायझेशन माध्यमातून टेस्टिंग करणे नियमाबाह्य असल्यामुळे आपण संबंधित केलेल्या संदर्भात मागील निविदेच्या अटी व शर्तीप्रमाणे NABL टेस्टिंग रिपोर्ट मान्य निविदा पास/नापास करण्यात यावे.</p> <p>३. शासन निर्णय दर दाखलावा/२०२४/प्र.क्र.३९/उद्योग-६ दि.१९/०३/२०२४ अन्वये २५ टक्के खरेदी ही अनुसूचित जाती/जमाती या उद्योजकांकडून आरक्षित केल्याबाबत वरील निविदेमध्ये त्याची अंमलबजावणी करणे आवश्यक आहे.</p>		<p>1) No change 2) No change 3) Govt. GR 01.12.2016 प्रमाणे कार्यवाही करण्यात येईल.</p>
9	M/s Brisk Surgicals Cotton Industries				<p>Price Matching Clause We note that the Price Matching concept is not incorporated in the tender terms and conditions. We kindly request you to consider incorporating the same in the tender.</p>		No change
					<p>Annexure X – Compliance Sheet for Technical Proposal Sr. No. 1: Absorbent Cotton Wool IP – 500 gms Kindly clarify whether the 500 gms refers to Net Weight or Gross Weight per packet.</p>	<p>Annexure X – Compliance Sheet for Technical Proposal Sr. No. 1: Absorbent Cotton Wool IP – 500 gms Kindly clarify whether the 500 gms refers to Net Weight or Gross Weight per packet.</p>	Absorbent Cotton Wool 500 gms IP

**E-253 Pre-Bid Query (General Queries)**

Sr. No.	Name of the Bidder	Tender Pg. No.	Tender Clause No.	Clause Title / Terms & conditions	Query / Clarification	Amendment Request	Department Remark
10	Narmada surgicals	12	<b>2. Eligibility criteria:</b>	<b>8) Average Annual Turnover</b> Average Annual Turnover (in last three financial years (2022-23,2023-24, 2024-25) (as attached Annexure-6)	1. As per Clause No.4.4.1.2 and 5.14 of Maharashtra Govt. G.R.Dt.01/12/2016 it is stated that 'no such clauses should not be mentioned in the terms and conditions of the tender which hamper competition and restrict eligibility of bidders' so keeping in mind that these items are now reserved. We request you to reduce the Turnover criteria for MSME. HBPCL also kept Rs. One Crore. We once again request you to grant exemption for the same.	1. As per Clause No.4.4.1.2 and 5.14 of Maharashtra Govt. G.R.Dt.01/12/2016 it is stated that 'no such clauses should not be mentioned in the terms and conditions of the tender which hamper competition and restrict eligibility of bidders' so keeping in mind that these items are now reserved. We request you to reduce the Turnover criteria for MSME. HBPCL also kept Rs. One Crore. We once again request you to grant exemption for the same.	No change
10	Narmada surgicals	12	<b>2. Eligibility criteria:</b>	<b>10) Past Work Experience</b> The bidder must submit particulars of quantity of the past supplies made as per Annex 3. Out of this at least 25 % quantity for "Similar item as specified in the Technical Specification and in the Schedule of Requirements" must have been supplied in any of the last 3 (Three) Financial years i.e. 2022-23, 2023-24,2024-25	2. As per T&C for 'Past Work Experience' the bidder must submit production sale data of at least 25% in any of the past three years along with order copies. We request you to amend this condition of supply as this condition relaxed to 10% for MSME.	2. As per T&C for 'Past Work Experience' the bidder must submit production sale data of at least 25% in any of the past three years along with order copies. We request you to amend this condition of supply as this condition relaxed to 10% for MSME.	No change

				
<b>Pharmacist</b>	<b>Senior. Pharmacy Officer</b>	<b>Technical Officer (Technical)</b>	<b>Chief Acc. &amp; Finance Officer</b>	<b>General Manager (Technical)</b>
Miss. Ojasvi Devlekar	Mr. Dnyaneshwar Visawe	Dr. D. B. Kheikar	Mr. Bhagwat Murkute	Dr. Sanjeevkumar Jadhav
MMGPA, Mumbai	MMGPA, Mumbai	MMGPA, Mumbai	MMGPA, Mumbai	MMGPA, Mumbai