

Madhya Pradesh Public Health Services Corporation Ltd.

Pre-bid Query Resolution Sheet

Tender Ref. No. T-900/MPPHSCL/Surgical and Consumables /RC/2026, Dt. 12-05-2026

(Uploading Date: 20-05-2026)

Sr. No.	Item Code	Item Name	Specification in the Floated Tender	Recommendation from bidder with Justification	Amendment/Resolution
1	700890	MVA Kit (Manual Vacuum Aspiration Kit, Consumable	International Certification : Conforms to CE and ISO requirements or is GMP manufactured.	Conforms to European CE, ISO and USFDA certification requirements or is GMP manufactured. This will help ensure higher product quality, safety, and compliance with international standards. Conforms to European CE, ISO and USFDA certification requirements or is GMP manufactured.	No Change As per tender terms & conditions
2				Standard Cannula Specification Cannula should be autoclavable/reusable at least 25 times (same as aspirator) so that the complete set can be reused after proper sterilization This will allow repeated usage and will be beneficial for hospitals and end users. Cannula should be autoclavable/reusable at least 25 times (same as aspirator) so that the complete set can be reused after proper sterilization	No Change As per tender terms & Specification